



ITS-90 Table for type T thermocouple
Thermoelectric Voltage in mV

T Cu / Cu-Ni

°C	0	-1	-2	-3	-4	-5	-6	-7	-8	-9	-10
-270	-6.258										
-260	-6.232	-6.236	-6.239	-6.242	-6.245	-6.248	-6.251	-6.253	-6.255	-6.256	-6.258
-250	-6.18	-6.187	-6.193	-6.198	-6.204	-6.209	-6.214	-6.219	-6.223	-6.228	-6.232
-240	-6.105	-6.114	-6.122	-6.13	-6.138	-6.146	-6.153	-6.16	-6.167	-6.174	-6.18
-230	-6.007	-6.017	-6.028	-6.038	-6.049	-6.059	-6.068	-6.078	-6.087	-6.096	-6.105
-220	-5.888	-5.901	-5.914	-5.926	-5.938	-5.95	-5.962	-5.973	-5.985	-5.996	-6.007
-210	-5.753	-5.767	-5.782	-5.795	-5.809	-5.823	-5.836	-5.85	-5.863	-5.876	-5.888
-200	-5.603	-5.619	-5.634	-5.65	-5.665	-5.68	-5.695	-5.71	-5.724	-5.739	-5.753
-190	-5.439	-5.456	-5.473	-5.489	-5.506	-5.523	-5.539	-5.555	-5.571	-5.587	-5.603
-180	-5.261	-5.279	-5.297	-5.316	-5.334	-5.351	-5.369	-5.387	-5.404	-5.421	-5.439
-170	-5.07	-5.089	-5.109	-5.128	-5.148	-5.167	-5.186	-5.205	-5.224	-5.242	-5.261
-160	-4.865	-4.886	-4.907	-4.928	-4.949	-4.969	-4.989	-5.01	-5.03	-5.05	-5.07
-150	-4.648	-4.671	-4.693	-4.715	-4.737	-4.759	-4.78	-4.802	-4.823	-4.844	-4.865
-140	-4.419	-4.443	-4.466	-4.489	-4.512	-4.535	-4.558	-4.581	-4.604	-4.626	-4.648
-130	-4.177	-4.202	-4.226	-4.251	-4.275	-4.3	-4.324	-4.348	-4.372	-4.395	-4.419
-120	-3.923	-3.949	-3.975	-4	-4.026	-4.052	-4.077	-4.102	-4.127	-4.152	-4.177
-110	-3.657	-3.684	-3.711	-3.738	-3.765	-3.791	-3.818	-3.844	-3.871	-3.897	-3.923
-100	-3.379	-3.407	-3.435	-3.463	-3.491	-3.519	-3.547	-3.574	-3.602	-3.629	-3.657
-90	-3.089	-3.118	-3.148	-3.177	-3.206	-3.235	-3.264	-3.293	-3.322	-3.35	-3.379
-80	-2.788	-2.818	-2.849	-2.879	-2.91	-2.94	-2.97	-3	-3.03	-3.059	-3.089
-70	-2.476	-2.507	-2.539	-2.571	-2.602	-2.633	-2.664	-2.695	-2.726	-2.757	-2.788
-60	-2.153	-2.186	-2.218	-2.251	-2.283	-2.316	-2.348	-2.38	-2.412	-2.444	-2.476
-50	-1.819	-1.853	-1.887	-1.92	-1.954	-1.987	-2.021	-2.054	-2.087	-2.12	-2.153
-40	-1.475	-1.51	-1.545	-1.579	-1.614	-1.648	-1.683	-1.717	-1.751	-1.785	-1.819
-30	-1.121	-1.157	-1.192	-1.228	-1.264	-1.299	-1.335	-1.37	-1.405	-1.44	-1.475
-20	-0.757	-0.794	-0.83	-0.867	-0.904	-0.94	-0.976	-1.013	-1.049	-1.085	-1.121
-10	-0.383	-0.421	-0.459	-0.496	-0.534	-0.571	-0.608	-0.646	-0.683	-0.72	-0.757
0	0	-0.039	-0.077	-0.116	-0.154	-0.193	-0.231	-0.269	-0.307	-0.345	-0.383
°C	0	-1	-2	-3	-4	-5	-6	-7	-8	-9	-10

°C	0	1	2	3	4	5	6	7	8	9	10
0	0	0.039	0.078	0.117	0.156	0.195	0.234	0.273	0.312	0.352	0.391
10	0.391	0.431	0.47	0.51	0.549	0.589	0.629	0.669	0.709	0.749	0.79
20	0.79	0.83	0.87	0.911	0.951	0.992	1.033	1.074	1.114	1.155	1.196
30	1.196	1.238	1.279	1.32	1.362	1.403	1.445	1.486	1.528	1.57	1.612
40	1.612	1.654	1.696	1.738	1.78	1.823	1.865	1.908	1.95	1.993	2.036
50	2.036	2.079	2.122	2.165	2.208	2.251	2.294	2.338	2.381	2.425	2.468
60	2.468	2.512	2.556	2.6	2.643	2.687	2.732	2.776	2.82	2.864	2.909
70	2.909	2.953	2.998	3.043	3.087	3.132	3.177	3.222	3.267	3.312	3.358
80	3.358	3.403	3.448	3.494	3.539	3.585	3.631	3.677	3.722	3.768	3.814
90	3.814	3.86	3.907	3.953	3.999	4.046	4.092	4.138	4.185	4.232	4.279
100	4.279	4.325	4.372	4.419	4.466	4.513	4.561	4.608	4.655	4.702	4.75
110	4.75	4.798	4.845	4.893	4.941	4.988	5.036	5.084	5.132	5.18	5.228
120	5.228	5.277	5.325	5.373	5.422	5.47	5.519	5.567	5.616	5.665	5.714
130	5.714	5.763	5.812	5.861	5.91	5.959	6.008	6.057	6.107	6.156	6.206
140	6.206	6.255	6.305	6.355	6.404	6.454	6.504	6.554	6.604	6.654	6.704
150	6.704	6.754	6.805	6.855	6.905	6.956	7.006	7.057	7.107	7.158	7.209
160	7.209	7.26	7.31	7.361	7.412	7.463	7.515	7.566	7.617	7.668	7.72
°C	0	1	2	3	4	5	6	7	8	9	10



ITS-90 Table for type T thermocouple
Thermoelectric Voltage in mV

T Cu / Cu-Ni

°C	0	1	2	3	4	5	6	7	8	9	10
170	7.72	7.771	7.823	7.874	7.926	7.977	8.029	8.081	8.133	8.185	8.237
180	8.237	8.289	8.341	8.393	8.445	8.497	8.55	8.602	8.654	8.707	8.759
190	8.759	8.812	8.865	8.917	8.97	9.023	9.076	9.129	9.182	9.235	9.288
200	9.288	9.341	9.395	9.448	9.501	9.555	9.608	9.662	9.715	9.769	9.822
210	9.822	9.876	9.93	9.984	10.038	10.092	10.146	10.2	10.254	10.308	10.362
220	10.362	10.417	10.471	10.525	10.58	10.634	10.689	10.743	10.798	10.853	10.907
230	10.907	10.962	11.017	11.072	11.127	11.182	11.237	11.292	11.347	11.403	11.458
240	11.458	11.513	11.569	11.624	11.68	11.735	11.791	11.846	11.902	11.958	12.013
250	12.013	12.069	12.125	12.181	12.237	12.293	12.349	12.405	12.461	12.518	12.574
260	12.574	12.63	12.687	12.743	12.799	12.856	12.912	12.969	13.026	13.082	13.139
270	13.139	13.196	13.253	13.31	13.366	13.423	13.48	13.537	13.595	13.652	13.709
280	13.709	13.766	13.823	13.881	13.938	13.995	14.053	14.11	14.168	14.226	14.283
290	14.283	14.341	14.399	14.456	14.514	14.572	14.63	14.688	14.746	14.804	14.862
300	14.862	14.92	14.978	15.036	15.095	15.153	15.211	15.27	15.328	15.386	15.445
310	15.445	15.503	15.562	15.621	15.679	15.738	15.797	15.856	15.914	15.973	16.032
320	16.032	16.091	16.15	16.209	16.268	16.327	16.387	16.446	16.505	16.564	16.624
330	16.624	16.683	16.742	16.802	16.861	16.921	16.98	17.04	17.1	17.159	17.219
340	17.219	17.279	17.339	17.399	17.458	17.518	17.578	17.638	17.698	17.759	17.819
350	17.819	17.879	17.939	17.999	18.06	18.12	18.18	18.241	18.301	18.362	18.422
360	18.422	18.483	18.543	18.604	18.665	18.725	18.786	18.847	18.908	18.969	19.03
370	19.03	19.091	19.152	19.213	19.274	19.335	19.396	19.457	19.518	19.579	19.641
380	19.641	19.702	19.763	19.825	19.886	19.947	20.009	20.07	20.132	20.193	20.255
390	20.255	20.317	20.378	20.44	20.502	20.563	20.625	20.687	20.748	20.81	20.872
400	20.872										
°C	0	1	2	3	4	5	6	7	8	9	10